

RESISTIVITY VERIFICATION SAMPLE
VERIFICATION CHART

Sample #: MNS-CALSIG-L-80K-30
SIG Model #: SIG2010
SIG Serial #: SIG2010-1
Measured by:
Date:
Location: Mecanum Inc.



Diameter: 119.8 mm
Test Diameter : 100 mm
Hole Diameter : 3.00 ± 0.01 mm
Thickness : 30 mm

Expected Resistivity @ 100 CCM: **83700 ± 1200 MKS rayls/m or Ns/m^4**

Expected Static Resistivity @ 0 CCM: **72544 ± 1000 MKS rayls/m or Ns/m^4**

Signed: _____